

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.089 °C	-25.6 °C	-0.511 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.121 °C	-0.4 °C	-0.279 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	4.6 °C	-0.393 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.0 °C	-0.007 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	24.9 °C	-0.449 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	29.7 °C	-0.362 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.190 °C	40.0 °C	-0.190 °C	± 0.5 °C	± 4.0E-01 °C

