

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.089 °C	-25.3 °C	-0.211 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.121 °C	-0.2 °C	-0.079 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.1 °C	0.107 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	10.2 °C	0.193 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.349 °C	25.1 °C	-0.249 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	29.9 °C	-0.162 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.190 °C	39.7 °C	-0.490 °C	± 0.5 °C	± 4.0E-01 °C

