

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.089 °C        | -25.2 °C  | -0.111 °C     | ± 0.7 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.121 °C         | -0.1 °C   | 0.021 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 4.993 °C          | 5.3 °C    | 0.307 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.007 °C         | 10.0 °C   | -0.007 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.349 °C         | 25.2 °C   | -0.149 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.062 °C         | 30.0 °C   | -0.062 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.190 °C         | 40.1 °C   | -0.090 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

