

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.089 °C	-25.3 °C	-0.211 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.722 °C	-0.6 °C	0.122 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.495 °C	5.4 °C	-0.095 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.308 °C	10.0 °C	-0.308 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.7 °C	-0.349 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.163 °C	29.9 °C	-0.263 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.491 °C	40.2 °C	-0.291 °C	± 0.5 °C	± 4.0E-01 °C

