

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.089 °C	-24.7 °C	0.389 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.722 °C	-0.3 °C	0.422 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.495 °C	5.7 °C	0.205 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.308 °C	10.1 °C	-0.208 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.7 °C	-0.349 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.163 °C	30.0 °C	-0.163 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.491 °C	40.2 °C	-0.291 °C	± 0.5 °C	± 4.0E-01 °C

