

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.089 °C        | -24.7 °C  | 0.389 °C      | ± 0.7 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.722 °C         | -0.4 °C   | 0.322 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 5.495 °C          | 5.8 °C    | 0.305 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 10.308 °C         | 10.0 °C   | -0.308 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 25.049 °C         | 25.0 °C   | -0.049 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.163 °C         | 30.2 °C   | 0.037 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.491 °C         | 40.7 °C   | 0.209 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |

