

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.322 °C	14.8 °C	-0.522 °C	± 1.0 °C	± 1.8E-01 °C
20 °C	20.336 °C	19.9 °C	-0.436 °C	± 1.0 °C	± 1.8E-01 °C
23 °C	23.544 °C	22.7 °C	-0.844 °C	± 1.0 °C	± 1.8E-01 °C
30 °C	30.062 °C	30.3 °C	0.238 °C	± 1.0 °C	± 1.6E-01 °C
30 %HR	30.570 %HR	24.9 %HR	-5.670 %HR	± 5.0 %HR	± 1.2E+00 %HR
50 %HR	50.110 %HR	43.4 %HR	-6.710 %HR	± 5.0 %HR	± 1.2E+00 %HR
70 %HR	70.448 %HR	62.3 %HR	-8.148 %HR	± 5.0 %HR	± 1.2E+00 %HR
75 %HR	75.560 %HR	65.3 %HR	-10.260 %HR	± 5.0 %HR	± 1.2E+00 %HR