

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | E.M.P <i>(Tolerance)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|--------------------------|------------------------------------|------------|
| 23.3 μm                  | 22 μm            | -0.97 μm             | ± 10 μm                  | ±                                  | 8.2E-01 μm |
| 51.6 μm                  | 51 μm            | -0.27 μm             | ± 10 μm                  | ±                                  | 8.2E-01 μm |
| 124.6 μm                 | 125 μm           | 0.40 μm              | ± 10 μm                  | ±                                  | 5.8E-01 μm |
| 250.3 μm                 | 250 μm           | 0.03 μm              | ± 18 μm                  | ±                                  | 1.3E+00 μm |
| 486.0 μm                 | 485 μm           | -1.00 μm             | ± 25 μm                  | ±                                  | 5.8E-01 μm |
| 963.0 μm                 | 963 μm           | 0.00 μm              | ± 39 μm                  | ±                                  | 6.0E-01 μm |