

Certificado No. (Certificate #):		55730	Equipo (Instrument): INDICADOR DE DIAL		
Fecha de Calibración (Calibration Date):		2025-04-14	Marca (Brand): Wen		
Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0.0500	in	0.050 in	0.000 in	± in	± 5.8E-04 in
0.1010	in	0.328 in	0.227 in	± in	± 5.2E-01 in
0.1100	in	0.110 in	0.000 in	± in	± 5.8E-04 in
0.1500	in	0.488 in	0.338 in	± in	± 7.8E-01 in
0.2000	in	0.200 in	0.000 in	± in	± 5.9E-04 in
0.3000	in	0.300 in	0.000 in	± in	± 6.1E-04 in
0.4000	in	0.400 in	0.000 in	± in	± 6.4E-04 in
0.5000	in	0.500 in	0.000 in	± in	± 6.8E-04 in
1.0000	in	0.999 in	-0.001 in	± in	± 9.1E-04 in