

| Certificado No. (Certificate #):         |    | 55731      | Equipo (Instrument): INDICADOR DE DIAL |                      |                             |
|------------------------------------------|----|------------|----------------------------------------|----------------------|-----------------------------|
| Fecha de Calibración (Calibration Date): |    | 2025-04-14 | Marca (Brand): Wen                     |                      |                             |
| Patrón (Standard)                        |    | UBP (UUT)  | Error (Error)                          | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
| 0.0500                                   | in | 0.050      | in                                     | ± in                 | ± 5.8E-04 in                |
| 0.1010                                   | in | 0.101      | in                                     | ± in                 | ± 5.8E-04 in                |
| 0.1100                                   | in | 0.110      | in                                     | ± in                 | ± 5.8E-04 in                |
| 0.1500                                   | in | 0.150      | in                                     | ± in                 | ± 5.9E-04 in                |
| 0.2000                                   | in | 0.200      | in                                     | ± in                 | ± 5.9E-04 in                |
| 0.3000                                   | in | 0.300      | in                                     | ± in                 | ± 6.1E-04 in                |
| 0.4000                                   | in | 0.400      | in                                     | ± in                 | ± 6.4E-04 in                |
| 0.5000                                   | in | 0.500      | in                                     | ± in                 | ± 6.8E-04 in                |
| 1.0000                                   | in | 1.000      | in                                     | ± in                 | ± 9.1E-04 in                |