

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.007 °C	10.1 °C	0.093 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.770 °C	32.7 °C	-0.070 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.591 %HR	38.6 %HR	-1.991 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.253 %HR	73.3 %HR	-1.953 %HR	± 3.0 %HR	± 1.9E+00 %HR