

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 10 °C                      | 10.007 °C                  | 10.1 °C            | 0.093 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 32 °C                      | 32.770 °C                  | 32.7 °C            | -0.070 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 40 %HR                     | 40.591 %HR                 | 38.7 %HR           | -1.891 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 75 %HR                     | 75.253 %HR                 | 73.4 %HR           | -1.853 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |

