

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
75 mm	75.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
77 mm	77.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
78 mm	78.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
80 mm	80.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
81 mm	81.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
84 mm	84.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
88 mm	88.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
90 mm	90.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
92 mm	92.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
95 mm	95.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
100 mm	100.00 mm	0 µm	± 5 µm	±	5.8E+00 µm