

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
75 mm	75.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
77 mm	77.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
78 mm	78.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
80 mm	80.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
81 mm	81.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
84 mm	84.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
88 mm	88.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
90 mm	90.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
92 mm	92.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
95 mm	95.00 mm	0 μm	± 5 μm	±	5.8E+00 μm
100 mm	100.00 mm	0 μm	± 5 μm	±	5.8E+00 μm