

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
50 mm	50.00 mm	0 µm	± 4 µm	±	5.8E+00 µm
52 mm	52.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
55 mm	55.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
58 mm	58.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
60 mm	60.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
62 mm	62.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
65 mm	65.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
68 mm	68.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
70 mm	70.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
72 mm	72.00 mm	0 µm	± 5 µm	±	5.8E+00 µm
75 mm	75.00 mm	0 µm	± 5 µm	±	5.8E+00 µm