

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.00 mm	1.00 mm	0 µm	± 13 µm	±	#REF! µm
1.25 mm	1.26 mm	10 µm	± 13 µm	±	#REF! µm
1.50 mm	1.51 mm	10 µm	± 13 µm	±	#REF! µm
2.00 mm	2.01 mm	10 µm	± 13 µm	±	#REF! µm
3.00 mm	3.01 mm	10 µm	± 13 µm	±	#REF! µm
5.00 mm	5.01 mm	10 µm	± 13 µm	±	#REF! µm
6.00 mm	6.01 mm	10 µm	± 13 µm	±	#REF! µm
8.00 mm	8.02 mm	20 µm	± 13 µm	±	#REF! µm
10.00 mm	10.02 mm	20 µm	± 13 µm	±	#REF! µm