

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.273 °C	29.8 °C	-0.473 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.435 %HR	30 %HR	0.565 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.527 %HR	46 %HR	-3.527 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	74.752 %HR	65 %HR	-9.752 %HR	± 5.0 %HR	± 1.9E+00 %HR