

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.273 °C	30.1 °C	-0.173 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.435 %HR	25 %HR	-4.435 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.527 %HR	39 %HR	-10.527 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	74.752 %HR	63 %HR	-11.752 %HR	± 5.0 %HR	± 1.9E+00 %HR