

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.263 °C	29.4 °C	-0.863 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	33.0 %HR	-2.683 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.795 %HR	47.0 %HR	-3.795 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	69.0 %HR	-1.243 %HR	± 5.0 %HR	± 1.9E+00 %HR