

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 1	0.99985	0.150	± 1	± 6.6E-01 mm					
0 a 2	2.00059	-0.590	± 1.4	± 7.0E-01 mm					
0 a 3	3.00038	-0.380	± 1.8	± 7.0E-01 mm					
0 a 4	4.00117	-1.170	± 2.2	± 7.0E-01 mm					
0 a 5	4.99973	0.270	± 2.6	± 7.0E-01 mm					
0 a 6	5.99988	0.120	± 3	± 6.1E-01 mm					
0 a 7	6.99978	0.220	± 3.4	± 6.1E-01 mm					
0 a 8	8.00003	-0.030	± 3.8	± 6.1E-01 mm					
0 a 9	8.99995	0.050	± 4.2	± 6.1E-01 mm					
0 a 10	10.00027	-0.270	± 4.6	± 6.1E-01 mm					
0 a 11	11.00027	-0.270	± 5	± 6.1E-01 mm					
0 a 12	12.00063	-0.630	± 5.4	± 6.1E-01 mm					
0 a 13	12.99998	0.020	± 5.8	± 6.1E-01 mm					
0 a 14	14.00042	-0.420	± 6.2	± 6.1E-01 mm					
0 a 15	15.00013	-0.130	± 6.6	± 6.1E-01 mm					
0 a 16	16.00049	-0.490	± 7	± 6.1E-01 mm					
0 a 17	17.00049	-0.490	± 7.4	± 6.1E-01 mm					
0 a 18	18.00102	-1.020	± 7.8	± 6.1E-01 mm					
0 a 19	19.00060	-0.600	± 8.2	± 6.1E-01 mm					
0 a 20	20.00091	-0.910	± 8.6	± 6.1E-01 mm					
0 a 21	21.00048	-0.480	± 9	± 7.4E-01 mm					
0 a 22	22.00080	-0.800	± 9.4	± 7.4E-01 mm					
0 a 23	23.00059	-0.590	± 9.8	± 7.4E-01 mm					
0 a 24	24.00098	-0.980	± 10.2	± 7.4E-01 mm					
0 a 25	25.00065	-0.650	± 10.6	± 7.4E-01 mm					
0 a 26	26.00102	-1.020	± 11	± 7.4E-01 mm					
0 a 27	27.00068	-0.680	± 11.4	± 7.4E-01 mm					
0 a 28	28.00115	-1.150	± 11.8	± 7.4E-01 mm					
0 a 29	29.00071	-0.710	± 12.2	± 7.4E-01 mm					
0 a 30	30.00105	-1.050	± 12.6	± 7.4E-01 mm					