

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) |       | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|-------|--------------------------------------|
| 30 °C                      | 30.042 °C                  | 29.0 °C            | -1.042 °C              | ±                             | 1 °C  | ± 4.0E-01 °C                         |
| 35 %HR                     | 33.995 %HR                 | 38.0 %HR           | 4.005 %HR              | ±                             | 5 %HR | ± 1.9E+00 %HR                        |
| 50 %HR                     | 49.323 %HR                 | 53.0 %HR           | 3.677 %HR              | ±                             | 5 %HR | ± 1.9E+00 %HR                        |
| 70 %HR                     | 73.362 %HR                 | 78.0 %HR           | 4.638 %HR              | ±                             | 5 %HR | ± 1.9E+00 %HR                        |