

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	30 mm	30.00 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	60.0 mm	60.00 mm	0 µm	± 30 µm	± 5.8E+00 µm
Profundidad / (Depth)	90 mm	90.00 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	120 mm	120.00 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	150 mm	150.00 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	180 mm	180.00 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	210 mm	210.00 mm	0 µm	± 40 µm	± 6.1E+00 µm
Profundidad / (Depth)	270 mm	270.00 mm	0 µm	± 40 µm	± 6.1E+00 µm
Profundidad / (Depth)	300 mm	300.00 mm	0 µm	± 40 µm	± 6.1E+00 µm