

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	30.1 °C	0.038 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	37.6 %HR	1.917 %HR	± 0.5 %HR	± 1.9E+00 %HR
50 %HR	50.550 %HR	50.6 %HR	0.050 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	68.7 %HR	-1.748 %HR	± 3.0 %HR	± 1.9E+00 %HR