

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.062 °C	30.1 °C	0.038 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	37.4 %HR	1.717 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.550 %HR	51.6 %HR	1.050 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	67.8 %HR	-2.648 %HR	± 3.0 %HR	± 1.9E+00 %HR