

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.213 °C	29.7 °C	-0.513 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.310 %HR	30 %HR	1.690 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	48.883 %HR	45 %HR	-3.883 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	76.102 %HR	75 %HR	-1.102 %HR	± 5.0 %HR	± 1.9E+00 %HR