

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.3 μm	23.4 μm	0.1 μm	± 2.2 μm	±	5.8E-02 μm
51.6 μm	51 μm	-0.6 μm	± 3.0 μm	±	5.8E-02 μm
124.6 μm	124 μm	-0.6 μm	± 5.2 μm	±	6.1E-02 μm
250.3 μm	249 μm	-1.3 μm	± 9.0 μm	±	6.9E-02 μm
486.0 μm	490 μm	4.0 μm	± 16 μm	±	9.4E-02 μm
963.0 μm	960 μm	-3.0 μm	± 30 μm	±	1.6E-01 μm