

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.213 °C	30.3 °C	0.087 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.648 %HR	31.0 %HR	2.352 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	48.883 %HR	51.0 %HR	2.117 %HR	± 2.5 %HR	± 1.9E+00 %HR
75 %HR	76.102 %HR	76.4 %HR	0.298 %HR	± 2.5 %HR	± 1.9E+00 %HR