

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.391 °C	0.0 °C	0.391 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.3 °C	0.279 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.6 °C	0.237 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	38.5 %HR	2.817 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.998 %HR	52.1 %HR	2.102 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.550 %HR	70.4 %HR	-0.150 %HR	± 3.0 %HR	± 1.9E+00 %HR