

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.391 °C	-0.6 °C	-0.209 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.1 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.6 °C	0.237 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	37.1 %HR	1.417 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.998 %HR	52.7 %HR	2.702 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.550 %HR	70.5 %HR	-0.050 %HR	± 3.0 %HR	± 1.9E+00 %HR