

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.391 °C	-0.5 °C	-0.109 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.2 °C	0.179 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.1 °C	-0.263 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	37.9 %HR	2.217 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.998 %HR	52.7 %HR	2.702 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.550 %HR	69.0 %HR	-1.550 %HR	± 3.0 %HR	± 1.9E+00 %HR