

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.985 °C	2.0 °C	0.015 °C	± 0.5 °C	± 4.0E-01 °C
4 °C	3.991 °C	4.0 °C	0.009 °C	± 0.5 °C	± 4.0E-01 °C
6 °C	5.996 °C	6.0 °C	0.004 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.002 °C	8.1 °C	0.098 °C	± 0.5 °C	± 4.0E-01 °C