

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.213 °C	30.4 °C	0.187 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.310 %HR	31.2 %HR	2.890 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	48.883 %HR	50.8 %HR	1.917 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	76.102 %HR	73.2 %HR	-2.902 %HR	± 3.0 %HR	± 1.9E+00 %HR