

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | E.M.P <i>(Tolerance)</i> | Incertidumbre <i>(Uncertainty)</i> | |
|--------------------------|------------------|----------------------|--------------------------|------------------------------------|-------------|
| 1 in | 1.0000 in | 0 in | ± 0.0001 in | ± | 7.60E-02 in |
| 2 in | 2.0000 in | 0 in | ± 0.0002 in | ± | 7.00E-02 in |
| 3 in | 3.0000 in | 0 in | ± 0.0002 in | ± | 7.60E-02 in |
| 4 in | 4.0000 in | 0 in | ± 0.0002 in | ± | 6.90E-02 in |
| 5 in | 5.0000 in | 0 in | ± 0.0002 in | ± | 1.1E-01 in |