

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1 in	1.0000 in	0 in	± 0.0001 in	±	7.60E-02 in
2 in	2.0000 in	0 in	± 0.0002 in	±	7.00E-02 in
3 in	3.0000 in	0 in	± 0.0002 in	±	7.60E-02 in
4 in	4.0000 in	0 in	± 0.0002 in	±	6.90E-02 in
5 in	5.0000 in	0 in	± 0.0002 in	±	1.1E-01 in
6 in	6.0000 in	0 in	± 0.0002 in	±	1.3E-01 in