

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
5.0000 in	5.0000 in	0 in	± 0.0002 in	±	7.0E-02 in
5.0625 in	5.0625 in	0 in	± 0.0002 in	±	5.7E-02 in
5.1000 in	5.1000 in	0 in	± 0.0002 in	±	7.6E-02 in
5.1250 in	5.1250 in	0 in	± 0.0002 in	±	9.4E-02 in
5.2000 in	5.2000 in	0 in	± 0.0002 in	±	7.0E-02 in
5.2500 in	5.2500 in	0 in	± 0.0002 in	±	9.9E-02 in
5.3000 in	5.3000 in	0 in	± 0.0002 in	±	7.6E-02 in
5.4000 in	5.4000 in	0 in	± 0.0002 in	±	9.4E-02 in
5.5000 in	5.5000 in	0 in	± 0.0002 in	±	6.9E-02 in
5.8000 in	5.8000 in	0 in	± 0.0002 in	±	1.1E-01 in
6.0000 in	6.0000 in	0 in	± 0.0002 in	±	1.3E-01 in