

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.0000 in	1.0000 in	0 in	± 0.0001 in	±	7.0E-02 in
1.0625 in	1.0625 in	0 in	± 0.0002 in	±	5.7E-02 in
1.1000 in	1.1000 in	0 in	± 0.0002 in	±	7.6E-02 in
1.1250 in	1.1250 in	0 in	± 0.0002 in	±	9.4E-02 in
1.2000 in	1.2000 in	0 in	± 0.0002 in	±	7.0E-02 in
1.2500 in	1.2500 in	0 in	± 0.0002 in	±	9.9E-02 in
1.3000 in	1.3000 in	0 in	± 0.0002 in	±	7.6E-02 in
1.4000 in	1.4000 in	0 in	± 0.0002 in	±	9.4E-02 in
1.5000 in	1.5000 in	0 in	± 0.0002 in	±	6.9E-02 in
1.8000 in	1.8000 in	0 in	± 0.0002 in	±	1.1E-01 in
2.0000 in	2.0000 in	0 in	± 0.0002 in	±	1.3E-01 in