

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 1	0.99959	0.410	± 1	± 6.6E-01 mm					
0 a 2	2.00005	-0.050	± 1.4	± 7.0E-01 mm					
0 a 3	3.00002	-0.020	± 1.8	± 7.0E-01 mm					
0 a 4	4.00023	-0.230	± 2.2	± 7.0E-01 mm					
0 a 5	4.99969	0.310	± 2.6	± 7.0E-01 mm					
0 a 6	6.00003	-0.030	± 3	± 6.1E-01 mm					
0 a 7	6.99963	0.370	± 3.4	± 6.1E-01 mm					
0 a 8	8.00000	0.000	± 3.8	± 6.1E-01 mm					
0 a 9	8.99972	0.280	± 4.2	± 6.1E-01 mm					
0 a 10	9.99991	0.090	± 4.6	± 6.1E-01 mm					
0 a 11	11.00015	-0.150	± 5	± 6.1E-01 mm					
0 a 12	12.00045	-0.450	± 5.4	± 6.1E-01 mm					
0 a 13	13.00003	-0.030	± 5.8	± 6.1E-01 mm					
0 a 14	14.00034	-0.340	± 6.2	± 6.1E-01 mm					
0 a 15	15.00017	-0.170	± 6.6	± 6.1E-01 mm					
0 a 16	16.00058	-0.580	± 7	± 6.1E-01 mm					
0 a 17	17.00016	-0.160	± 7.4	± 6.1E-01 mm					
0 a 18	18.00067	-0.670	± 7.8	± 6.1E-01 mm					
0 a 19	19.00060	-0.600	± 8.2	± 6.1E-01 mm					
0 a 20	20.00096	-0.960	± 8.6	± 6.1E-01 mm					
0 a 21	21.00047	-0.470	± 9	± 7.4E-01 mm					
0 a 22	22.00084	-0.840	± 9.4	± 7.4E-01 mm					
0 a 23	23.00039	-0.390	± 9.8	± 7.4E-01 mm					
0 a 24	24.00075	-0.750	± 10.2	± 7.4E-01 mm					
0 a 25	25.00050	-0.500	± 10.6	± 7.4E-01 mm					
0 a 26	26.00071	-0.710	± 11	± 7.4E-01 mm					
0 a 27	27.00059	-0.590	± 11.4	± 7.4E-01 mm					
0 a 28	28.00094	-0.940	± 11.8	± 7.4E-01 mm					
0 a 29	29.00070	-0.700	± 12.2	± 7.4E-01 mm					
0 a 30	30.00093	-0.930	± 12.6	± 7.4E-01 mm					