

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.629 °C	11.3 °C	0.671 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	19.965 °C	18.7 °C	-1.265 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.323 °C	29.2 °C	-1.123 °C	± 1.0 °C	± 4.0E-01 °C
40 °C	40.521 °C	38.3 °C	-2.221 °C	± 1.0 °C	± 4.0E-01 °C