

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.021 °C	14.7 °C	-0.321 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.650 °C	26.1 °C	0.450 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.363 °C	30.7 °C	0.337 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	70.550 %HR	71.1 %HR	0.550 %HR	± 3.0 %HR	± 1.9E+00 %HR