

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.085 °C	1.8 °C	-0.285 °C	± 1.0 °C	± 4.0E-01 °C
10 °C	10.408 °C	9.9 °C	-0.508 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	20.038 %HR	23.0 %HR	2.962 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	78.219 %HR	74.7 %HR	-3.559 %HR	± 5.0 %HR	± 1.9E+00 %HR