

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.021 °C	15.4 °C	0.379 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.650 °C	25.6 °C	-0.050 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	20.038 %HR	23.0 %HR	2.962 %HR	± 5.0 %HR	± 1.9E+00 %HR
65 %HR	65.540 %HR	61.0 %HR	-4.540 %HR	± 5.0 %HR	± 1.9E+00 %HR