

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.021 °C	14.3 °C	-0.721 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.650 °C	26.3 °C	0.650 °C	± 1.0 °C	± 4.0E-01 °C
20 %HR	20.038 %HR	20.0 %HR	-0.038 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.673 %HR	77.0 %HR	-3.673 %HR	± 5.0 %HR	± 1.9E+00 %HR