

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	1.985 °C	1.3 °C	-0.685 °C	± 1.0 °C	± 4.0E-01 °C
5 °C	4.993 °C	4.3 °C	-0.693 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	8.002 °C	7.3 °C	-0.702 °C	± 1.0 °C	± 4.0E-01 °C