

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 2 °C              | 1.985 °C          | 1.3 °C    | -0.685 °C     | ± 1.0 °C             | ± 4.0E-01 °C                |
| 5 °C              | 4.993 °C          | 4.2 °C    | -0.793 °C     | ± 1.0 °C             | ± 4.0E-01 °C                |
| 8 °C              | 8.002 °C          | 7.1 °C    | -0.902 °C     | ± 1.0 °C             | ± 4.0E-01 °C                |