

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
200 RPM	200 RPM	0 RPM	± 4 RPM	±	5.8E-01 RPM
300 RPM	300 RPM	0 RPM	± 6 RPM	±	5.8E-01 RPM
500 RPM	500 RPM	0 RPM	± 10 RPM	±	5.8E-01 RPM
1000 RPM	1000 RPM	0 RPM	± 20 RPM	±	5.8E-01 RPM
10000 RPM	10000 RPM	0 RPM	± 200 RPM	±	8.0E-01 RPM