

Tipo de Medida (<i>Measurement Type</i>)	Patrón (<i>Standard</i>)	UBP (<i>UT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)
Profundidad / (Depth)	1.00 mm	1.00 mm	0 μm	± 50 μm	± 5.8E+01 μm
Profundidad / (Depth)	1.50 mm	1.50 mm	0 μm	± 50 μm	± 5.8E+01 μm
Profundidad / (Depth)	3.00 mm	3.00 mm	0 μm	± 50 μm	± 5.8E+01 μm
Profundidad / (Depth)	5.00 mm	5.00 mm	0 μm	± 50 μm	± 5.8E+01 μm
Profundidad / (Depth)	10.00 mm	10.00 mm	0 μm	± 50 μm	± 5.8E+01 μm
Profundidad / (Depth)	15.00 mm	15.00 mm	0 μm	± 50 μm	± 5.8E+01 μm
Profundidad / (Depth)	20.00 mm	20.00 mm	0 μm	± 50 μm	± 5.8E+01 μm
Profundidad / (Depth)	25.00 mm	25.00 mm	0 μm	± 50 μm	± 5.8E+01 μm
Profundidad / (Depth)	30.00 mm	30.00 mm	0 μm	± 50 μm	± 5.8E+01 μm