

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	1.00 mm	1.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	1.50 mm	1.5 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	3.00 mm	3.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	5.00 mm	5.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	10.00 mm	10.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	15.00 mm	15.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	20.00 mm	20.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	25.00 mm	25.0 mm	0 µm	± 50 µm	± 5.8E+01 µm
Profundidad / (Depth)	30.00 mm	30.0 mm	0 µm	± 50 µm	± 5.8E+01 µm