

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.272 °C	15.1 °C	-0.172 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.299 °C	25.0 °C	-0.299 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.319 °C	32.5 °C	0.181 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.171 %HR	22.0 %HR	1.829 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.202 %HR	51.1 %HR	0.898 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.9 %HR	1.452 %HR	± 5.0 %HR	± 1.9E+00 %HR

