

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 15.272 °C                  | 15.3 °C            | 0.028 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.299 °C                  | 25.4 °C            | 0.101 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 32 °C                      | 32.319 °C                  | 32.5 °C            | 0.181 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 %HR                     | 20.171 %HR                 | 21.7 %HR           | 1.529 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.202 %HR                 | 50.9 %HR           | 0.698 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.448 %HR                 | 70.9 %HR           | 0.452 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |