

| Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | E.M.P <i>(Tolerance)</i> | Incertidumbre <i>(Uncertainty)</i> |            |
|--------------------------|------------------|----------------------|--------------------------|------------------------------------|------------|
| 52.65 μm                 | 52.70 μm         | 0.05 μm              | ± 2.5 μm                 | ±                                  | 5.8E-01 μm |
| 155.23 μm                | 154.6 μm         | -0.63 μm             | ± 2.5 μm                 | ±                                  | 5.8E-01 μm |
| 257.12 μm                | 257.2 μm         | 0.08 μm              | ± 2.5 μm                 | ±                                  | 5.8E-01 μm |
| 511.72 μm                | 511.8 μm         | 0.08 μm              | ± 2.5 μm                 | ±                                  | 5.8E-01 μm |